

Operator: Validation Manager
Process: Validation
Session Start: 22-11-2013 09:22:12
Session Stop: 22-11-2013 09:22:23
Session Name: EtO Validation
Session Text:
Vessel: EtO sterilizer
Product: Pharma Medical
Time Zone: UTC offset 01:00:00



Limit Report

Name: EtO Validation test
Description: Test of temperature distribution in 55 points having minimum temperature of 50°C using TrackSense loggers.

Input parameters

Criteria: Passed when 50,00 °C < Value
Triggering : All, Temperature
Data Series: LC 1, LC 2, LC 3, LC 4, LC 5, LC 6, LC 7, LC 8, LC 9, LC10, LC11, LC12, LC13, LC14, LC15, LC16, LC17, LC18, LC19, LC20, LC21, LC22, LC23, LC24, LC25, LC26, LC27, LC28, LC29, LC30, LC31, LC32, LC33, LC34, LC35, LC36, LC37, LC38, LC39, LC40, LC41, LC42, LC43, LC44, LC45, LC46, LC47, LC48, LC49, LC50, LC51, LC52, LC53, LC54, LC55

Range

Time: 12-03-1997 19:00:00 to 13-03-1997 11:39:00

Triggering Interval

| from | to |
|---------------------|---------------------|
| 12-03-1997 21:58:00 | 13-03-1997 05:25:00 |

Summary Time: 07:27:00